

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030719\
 Data File : VX007906.D
 Acq On : 08 Mar 2019 02:04
 Operator : JC/SP
 Sample : K1753-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z-2-7

Quant Time: Mar 08 08:12:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	273159	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	433533	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	409245	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	196010	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	153991	53.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.88%	
35) Dibromofluoromethane	5.51	113	141630	51.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.14%	
50) Toluene-d8	8.71	98	529895	52.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.06%	
62) 4-Bromofluorobenzene	11.13	95	191578	53.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.30%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	9496	5.693	ug/l	94
18) Methyl Acetate	2.78	43	5350	1.510	ug/l	94
20) Methylene Chloride	2.86	84	106511	40.126	ug/l	98
25) 2-Butanone	4.70	43	5995	2.539	ug/l	93
43) Isopropyl Acetate	6.46	43	17472	2.534	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030719\
Data File : VX007906.D
Acq On : 08 Mar 2019 02:04
Operator : JC/SP
Sample : K1753-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z-2-7

Quant Time: Mar 08 08:12:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

