

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030719\
 Data File : VX007909.D
 Acq On : 08 Mar 2019 03:23
 Operator : JC/SP
 Sample : K1757-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-6AB-S

Quant Time: Mar 08 08:19:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	272833	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	432069	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	414038	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	198100	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	153485	53.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.66%	
35) Dibromofluoromethane	5.51	113	141495	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
50) Toluene-d8	8.71	98	533382	52.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.10%	
62) 4-Bromofluorobenzene	11.13	95	194473	54.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.30%	
Target Compounds						
16) Acetone	2.45	43	18554	11.138	ug/l	92
20) Methylene Chloride	2.85	84	22529786	8497.766	ug/l #	81
25) 2-Butanone	4.70	43	4447	1.886	ug/l	93
43) Isopropyl Acetate	6.46	43	16315	2.374	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030719\
Data File : VX007909.D
Acq On : 08 Mar 2019 03:23
Operator : JC/SP
Sample : K1757-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-6AB-S

Quant Time: Mar 08 08:19:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

