

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030719\
 Data File : VX007910.D
 Acq On : 08 Mar 2019 03:50
 Operator : JC/SP
 Sample : K1757-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-6AB-D

Quant Time: Mar 08 08:21:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	281096	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	437443	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	426182	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	205367	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	155224	52.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.70%	
35) Dibromofluoromethane	5.50	113	143499	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
50) Toluene-d8	8.71	98	547942	53.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.64%	
62) 4-Bromofluorobenzene	11.13	95	200965	55.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.56%	
Target Compounds						
16) Acetone	2.45	43	24302	14.159	ug/l	100
18) Methyl Acetate	2.78	43	9110	2.498	ug/l	98
20) Methylene Chloride	2.86	84	36528	13.373	ug/l	98
25) 2-Butanone	4.70	43	7744	3.187	ug/l	91
43) Isopropyl Acetate	6.46	43	16804	2.416	ug/l	99
95) Naphthalene	13.83	128	28171	2.080	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030719\
Data File : VX007910.D
Acq On : 08 Mar 2019 03:50
Operator : JC/SP
Sample : K1757-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-6AB-D

Quant Time: Mar 08 08:21:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

