

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030721\
 Data File : VX021023.D
 Acq On : 05 Mar 2021 10:58
 Operator : JC/MD
 Sample : VX0307MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0307MBL01

Quant Time: Mar 05 23:55:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	224646	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.823	114	407611	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	399869	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	177716	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.023	65	145646	48.00	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.00%
35) Dibromofluoromethane	5.464	113	127559	51.51	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.02%
50) Toluene-d8	8.694	98	516756	53.51	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.02%
62) 4-Bromofluorobenzene	11.118	95	198159	55.50	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	111.00%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030721\
Data File : VX021023.D
Acq On : 05 Mar 2021 10:58
Operator : JC/MD
Sample : VX0307MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0307MBL01

Quant Time: Mar 05 23:55:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

