

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030721\
 Data File : VX021033.D
 Acq On : 05 Mar 2021 14:43
 Operator : JC/MD
 Sample : M1584-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW6-GW-758-760

Quant Time: Mar 05 23:58:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	214205	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.829	114	405921	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	388522	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	170826	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	143917	49.74	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.48%
35) Dibromofluoromethane	5.470	113	123436	50.05	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.10%
50) Toluene-d8	8.694	98	498692	51.86	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.72%
62) 4-Bromofluorobenzene	11.118	95	193096	54.31	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.62%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.370	101	2976	1.29	ug/l	99
12) 1,1-Dichloroethene	2.358	96	916	0.38	ug/l	92
16) Acetone	2.417	43	3962	3.32	ug/l	100
27) cis-1,2-Dichloroethene	4.558	96	908	0.29	ug/l	97
30) Chloroform	5.188	83	3584	0.77	ug/l	100
38) Carbon Tetrachloride	5.753	117	3767	1.01	ug/l	93
44) Trichloroethene	7.194	130	63347	20.83	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Mar 05 23:58:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

