

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030721\
 Data File : VX021034.D
 Acq On : 05 Mar 2021 15:04
 Operator : JC/MD
 Sample : M1584-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW6-GW-778-780

Quant Time: Mar 05 23:58:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	215210	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	407759	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	396072	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	160107	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	146384	50.36	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.72%	
35) Dibromofluoromethane	5.470	113	126341	51.00	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.00%	
50) Toluene-d8	8.694	98	511642	52.97	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.94%	
62) 4-Bromofluorobenzene	11.118	95	185279	51.87	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 103.74%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.187	85	1126	0.48	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.370	101	13293	5.74	ug/l	97
12) 1,1-Dichloroethene	2.358	96	3017	1.25	ug/l	89
16) Acetone	2.422	43	3982	3.32	ug/l	90
27) cis-1,2-Dichloroethene	4.564	96	2260	0.72	ug/l	93
30) Chloroform	5.182	83	5011	1.07	ug/l	98
38) Carbon Tetrachloride	5.758	117	7310	1.95	ug/l	94
44) Trichloroethene	7.188	130	353574	115.76	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030721\
Data File : VX021034.D
Acq On : 05 Mar 2021 15:04
Operator : JC/MD
Sample : M1584-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW6-GW-778-780

Quant Time: Mar 05 23:58:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

