

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030721\
 Data File : VX021035.D
 Acq On : 05 Mar 2021 15:25
 Operator : JC/MD
 Sample : M1584-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW6-GW-798-800

Quant Time: Mar 05 23:59:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	220947	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.835	114	417761	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	403328	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	174914	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	150613	50.47	ug/l	0.00
Spiked Amount 50.000	Range	61 - 141	Recovery	= 100.94%		
35) Dibromofluoromethane	5.464	113	128678	50.70	ug/l	0.00
Spiked Amount 50.000	Range	69 - 133	Recovery	= 101.40%		
50) Toluene-d8	8.694	98	525026	53.05	ug/l	0.00
Spiked Amount 50.000	Range	65 - 126	Recovery	= 106.10%		
62) 4-Bromofluorobenzene	11.118	95	193209	52.80	ug/l	0.00
Spiked Amount 50.000	Range	58 - 135	Recovery	= 105.60%		
Target Compounds						Qvalue
16) Acetone	2.417	43	4224	3.43	ug/l	94
44) Trichloroethene	7.188	130	22909	7.32	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030721\
Data File : VX021035.D
Acq On : 05 Mar 2021 15:25
Operator : JC/MD
Sample : M1584-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW6-GW-798-800

Quant Time: Mar 05 23:59:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

