

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030721\
 Data File : VX021037.D
 Acq On : 05 Mar 2021 16:07
 Operator : JC/MD
 Sample : M1584-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW6-GW-818-820

Quant Time: Mar 05 23:59:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	223924	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	414429	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	395383	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	169958	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	149909	49.56	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.12%
35) Dibromofluoromethane	5.464	113	128976	51.22	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.44%
50) Toluene-d8	8.694	98	525630	53.54	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.08%
62) 4-Bromofluorobenzene	11.118	95	185000	50.96	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.92%
Target Compounds						
16) Acetone	2.417	43	2970	2.38	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030721\
Data File : VX021037.D
Acq On : 05 Mar 2021 16:07
Operator : JC/MD
Sample : M1584-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW6-GW-818-820

Quant Time: Mar 05 23:59:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

