

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030721\
 Data File : VX021039.D
 Acq On : 05 Mar 2021 16:50
 Operator : JC/MD
 Sample : M1584-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW6-GW-858-860

Quant Time: Mar 06 00:00:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	215561	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	392034	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	380751	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	163539	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	140949	48.41	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.82%
35) Dibromofluoromethane	5.470	113	122460	51.41	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.82%
50) Toluene-d8	8.694	98	491356	52.91	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.82%
62) 4-Bromofluorobenzene	11.118	95	184419	53.70	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.40%
Target Compounds						
16) Acetone	2.417	43	2468	2.06	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030721\
Data File : VX021039.D
Acq On : 05 Mar 2021 16:50
Operator : JC/MD
Sample : M1584-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW6-GW-858-860

Quant Time: Mar 06 00:00:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

