

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030721\
 Data File : VX021042.D
 Acq On : 05 Mar 2021 17:57
 Operator : JC/MD
 Sample : M1575-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IDW-GW805-AQ-030321

Quant Time: Mar 06 00:01:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	210834	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	389973	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	381795	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	165899	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	139271	48.90	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.80%
35) Dibromofluoromethane	5.464	113	121050	51.09	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.18%
50) Toluene-d8	8.694	98	491655	53.22	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.44%
62) 4-Bromofluorobenzene	11.118	95	184504	54.01	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.02%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.364	96	814	0.34	ug/l	96
16) Acetone	2.417	43	12510	10.66	ug/l	97
20) Methylene Chloride	2.835	84	2099	0.75	ug/l	89
25) 2-Butanone	4.635	43	3216	1.80	ug/l	93
27) cis-1,2-Dichloroethene	4.570	96	1083	0.35	ug/l	79
30) Chloroform	5.176	83	51386	11.17	ug/l	96
44) Trichloroethene	7.188	130	783	0.27	ug/l	68
51) 4-Methyl-2-Pentanone	8.618	43	4581	1.31	ug/l #	78
52) Toluene	8.765	92	2120	0.31	ug/l	99
68) m/p-Xylenes	10.341	106	1714	0.32	ug/l	99
84) 1,2,4-Trimethylbenzene	11.789	105	2950	0.27	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030721\
Data File : VX021042.D
Acq On : 05 Mar 2021 17:57
Operator : JC/MD
Sample : M1575-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IDW-GW805-AQ-030321

Quant Time: Mar 06 00:01:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

