

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030722\
 Data File : VX027325.D
 Acq On : 07 Mar 2022 14:32
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 08 07:37:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030722W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 07 14:49:16 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	91004	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	128585	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122940	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56704	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	60681	56.192	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.380%
35) Dibromofluoromethane	5.391	113	46467	55.139	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.280%
50) Toluene-d8	8.653	98	181782	59.128	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	118.260%#
62) 4-Bromofluorobenzene	11.085	95	55246	45.836	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.680%
Target Compounds						Qvalue
10) Methyl Iodide	2.459	142	604	3.646	ug/l #	88
17) Carbon Disulfide	2.514	76	3724	1.595	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030722\
Data File : VX027325.D
Acq On : 07 Mar 2022 14:32
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 08 07:37:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030722W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 07 14:49:16 2022
Response via : Initial Calibration

