


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030722\
 Data File : VX027316.D
 Acq On : 07 Mar 2022 09:43
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030722W.M
 Title : SW846 8260
 Last Update : Mon Mar 07 14:49:16 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.7	3337	PASS
75	95	30	60	50.3	9497	PASS
95	95	100	100	100.0	18879	PASS
96	95	5	9	6.9	1296	PASS
173	174	0.00	2	1.4	202	PASS
174	95	50	100	77.9	14702	PASS
175	174	5	9	7.4	1095	PASS
176	174	95	101	96.4	14180	PASS
177	176	5	9	6.4	906	PASS