

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030723\
 Data File : VX034359.D
 Acq On : 07 Mar 2023 09:44
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Mar 08 00:59:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	304842	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	500792	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	463756	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	235275	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	209091	50.929	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.860%	
35) Dibromofluoromethane	5.379	113	168384	51.641	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.280%	
50) Toluene-d8	8.647	98	599631	51.221	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.440%	
62) 4-Bromofluorobenzene	11.079	95	252899	56.477	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 112.960%	

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	1.167	85	127530	48.414	ug/l	100
3) Chloromethane	1.301	50	179251	47.287	ug/l	96
4) Vinyl Chloride	1.374	62	174546	50.560	ug/l	96
5) Bromomethane	1.599	94	94737	75.790	ug/l	99
6) Chloroethane	1.685	64	102703	59.176	ug/l	99
7) Trichlorofluoromethane	1.886	101	266872	57.140	ug/l	94
8) Diethyl Ether	2.130	74	101517	53.952	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.331	101	159758	48.571	ug/l	95
10) Methyl Iodide	2.453	142	198838	49.650	ug/l	95
11) Tert butyl alcohol	2.953	59	198662	216.349	ug/l #	84
12) 1,1-Dichloroethene	2.319	96	151651	45.994	ug/l	97
13) Acrolein	2.233	56	124299	296.541	ug/l	99
14) Allyl chloride	2.660	41	313872	44.876	ug/l	91
15) Acrylonitrile	3.063	53	481714	240.474	ug/l	99
16) Acetone	2.374	43	529575	250.057	ug/l	95
17) Carbon Disulfide	2.514	76	411279	43.385	ug/l	99
18) Methyl Acetate	2.703	43	318428	48.592	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	586002	50.356	ug/l	99
20) Methylene Chloride	2.788	84	181816	47.647	ug/l	98
21) trans-1,2-Dichloroethene	3.093	96	168152	48.669	ug/l	98
22) Diisopropyl ether	3.758	45	647527	52.809	ug/l	94
23) Vinyl Acetate	3.721	43	2517254	256.236	ug/l	100
24) 1,1-Dichloroethane	3.611	63	331401	49.232	ug/l	99
25) 2-Butanone	4.550	43	743139	242.515	ug/l	99
26) 2,2-Dichloropropane	4.471	77	288754	47.353	ug/l	98
27) cis-1,2-Dichloroethene	4.483	96	205558	50.242	ug/l	94
28) Bromochloromethane	4.891	49	151883	50.579	ug/l	87
29) Tetrahydrofuran	5.001	42	447554	239.416	ug/l	99
30) Chloroform	5.087	83	336627	52.079	ug/l	98
31) Cyclohexane	5.471	56	298161	48.643	ug/l	94
32) 1,1,1-Trichloroethane	5.379	97	290985	50.219	ug/l	99
36) 1,1-Dichloropropene	5.690	75	245934	51.001	ug/l	98
37) Ethyl Acetate	4.709	43	271724	50.420	ug/l	99
38) Carbon Tetrachloride	5.678	117	252315	50.848	ug/l	97
39) Methylcyclohexane	7.379	83	304112	49.006	ug/l	100
40) Benzene	6.032	78	728782	50.998	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030723\
 Data File : VX034359.D
 Acq On : 07 Mar 2023 09:44
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Mar 08 00:59:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	153540	51.135	ug/l	99
42) 1,2-Dichloroethane	6.086	62	280709	55.851	ug/l	97
43) Isopropyl Acetate	6.336	43	484851	50.982	ug/l	99
44) Trichloroethene	7.123	130	188005	51.267	ug/l	94
45) 1,2-Dichloropropane	7.428	63	196758	51.721	ug/l	98
46) Dibromomethane	7.580	93	133248	53.222	ug/l #	86
47) Bromodichloromethane	7.818	83	268711	52.299	ug/l	98
48) Methyl methacrylate	7.690	41	231323	51.247	ug/l	96
49) 1,4-Dioxane	7.653	88	89264	991.449	ug/l #	98
51) 4-Methyl-2-Pentanone	8.568	43	1406731	268.855	ug/l	99
52) Toluene	8.714	92	464791	53.678	ug/l	98
53) t-1,3-Dichloropropene	8.976	75	315834	53.403	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	326355	51.545	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	193982	55.501	ug/l	98
56) Ethyl methacrylate	9.116	69	323134	54.652	ug/l	96
57) 1,3-Dichloropropane	9.305	76	335508	55.174	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	667640	260.668	ug/l	95
59) 2-Hexanone	9.427	43	1102831	276.567	ug/l	97
60) Dibromochloromethane	9.519	129	211511	54.769	ug/l	100
61) 1,2-Dibromoethane	9.610	107	205907	55.705	ug/l	99
64) Tetrachloroethene	9.269	164	160866	49.148	ug/l	93
65) Chlorobenzene	10.080	112	498920	51.611	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	193961	51.829	ug/l	98
67) Ethyl Benzene	10.195	91	902797	51.983	ug/l	99
68) m/p-Xylenes	10.299	106	685691	103.170	ug/l	93
69) o-Xylene	10.640	106	344004	51.991	ug/l	94
70) Styrene	10.653	104	574183	53.357	ug/l	95
71) Bromoform	10.799	173	156595	51.264	ug/l	99
73) Isopropylbenzene	10.964	105	897822	50.071	ug/l	99
74) N-amyl acetate	10.842	43	442338	49.103	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.214	83	299132	48.775	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	347709	58.733	ug/l	82
77) Bromobenzene	11.195	156	216851	50.254	ug/l	83
78) n-propylbenzene	11.305	91	1059776	50.819	ug/l	98
79) 2-Chlorotoluene	11.366	91	640039	50.053	ug/l	95
80) 1,3,5-Trimethylbenzene	11.451	105	754042	50.018	ug/l	96
81) trans-1,4-Dichloro-2-b...	11.018	75	103528	44.923	ug/l	96
82) 4-Chlorotoluene	11.451	91	737398	50.630	ug/l	95
83) tert-Butylbenzene	11.713	119	753300	48.789	ug/l	92
84) 1,2,4-Trimethylbenzene	11.750	105	766973	50.317	ug/l	96
85) sec-Butylbenzene	11.890	105	943543	50.322	ug/l	97
86) p-Isopropyltoluene	12.006	119	780954	50.328	ug/l	96
87) 1,3-Dichlorobenzene	11.969	146	409605	50.615	ug/l	97
88) 1,4-Dichlorobenzene	12.043	146	409115	50.121	ug/l	97
89) n-Butylbenzene	12.335	91	704743	50.992	ug/l	98
90) Hexachloroethane	12.536	117	152880	47.121	ug/l	78
91) 1,2-Dichlorobenzene	12.335	146	397943	50.259	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	12.945	75	68259	46.674	ug/l	73
93) 1,2,4-Trichlorobenzene	13.585	180	249297	49.042	ug/l	97
94) Hexachlorobutadiene	13.725	225	102262	48.527	ug/l	99
95) Naphthalene	13.774	128	848109	49.526	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	251264	49.788	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030723\
Data File : VX034359.D
Acq On : 07 Mar 2023 09:44
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Quant Time: Mar 08 00:59:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050







Tgt Ion:168 Resp: 304842
Ion Ratio Lower Upper
168 100
99 57.6 39.6 59.4



#2
Dichlorodifluoromethane
Concen: 48.414 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 85 Resp: 127530
Ion Ratio Lower Upper
85 100
87 31.8 15.8 47.4









#3
Chloromethane
Concen: 47.287 ug/l
RT: 1.301 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 50 Resp: 179251
Ion Ratio Lower Upper
50 100
52 33.6 25.2 37.8



#4
Vinyl Chloride
Concen: 50.560 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 62 Resp: 174546
Ion Ratio Lower Upper
62 100
64 30.7 26.2 39.4









#5
Bromomethane
Concen: 75.790 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 94 Resp: 94737
Ion Ratio Lower Upper
94 100
96 94.1 74.3 111.5



#6
Chloroethane
Concen: 59.176 ug/l
RT: 1.685 min Scan# 98
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 64 Resp: 102703
Ion Ratio Lower Upper
64 100
66 32.0 26.1 39.1









#7
 Trichlorofluoromethane
 Concen: 57.140 ug/l
 RT: 1.886 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX034359.D
 Acq: 07 Mar 2023 09:44




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:101 Resp: 266872
 Ion Ratio Lower Upper
 101 100
 103 67.8 50.6 75.8



#8
 Diethyl Ether
 Concen: 53.952 ug/l
 RT: 2.130 min Scan# 171
 Delta R.T. 0.000 min
 Lab File: VX034359.D
 Acq: 07 Mar 2023 09:44

Tgt Ion: 74 Resp: 101517
 Ion Ratio Lower Upper
 74 100
 45 99.1 51.5 154.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 48.571 ug/l
RT: 2.331 min Scan# 204
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 159758
Ion Ratio Lower Upper
101 100
85 44.9 36.7 55.1
151 77.1 66.5 99.7



#10
Methyl Iodide
Concen: 49.650 ug/l
RT: 2.453 min Scan# 224
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion:142 Resp: 198838
Ion Ratio Lower Upper
142 100
127 42.5 37.3 55.9
141 14.2 11.5 17.3





#11

Tert butyl alcohol

Concen: 216.349 ug/l

RT: 2.953 min Scan# 306

Delta R.T. 0.000 min

Lab File: VX034359.D

Acq: 07 Mar 2023 09:44

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 59 Resp: 198662

Ion Ratio Lower Upper

59 100

57 9.8 3.7 5.5#



#12

1,1-Dichloroethene

Concen: 45.994 ug/l

RT: 2.319 min Scan# 202

Delta R.T. 0.000 min

Lab File: VX034359.D

Acq: 07 Mar 2023 09:44




Tgt Ion: 96 Resp: 151651


Ion Ratio Lower Upper

96 100

61 163.8 134.2 201.4

98 63.8 50.2 75.2









#13
Acrolein
Concen: 296.541 ug/l
RT: 2.233 min Scan# 188
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 124299
Ion Ratio Lower Upper
56 100
55 69.6 56.3 84.5



#14
Allyl chloride
Concen: 44.876 ug/l
RT: 2.660 min Scan# 258
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 41 Resp: 313872
Ion Ratio Lower Upper
41 100
39 64.0 59.0 88.4
76 32.0 24.6 37.0









#15
Acrylonitrile
Concen: 240.474 ug/l
RT: 3.063 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 53 Resp: 481714
Ion Ratio Lower Upper
53 100
52 82.6 65.7 98.5
51 35.9 29.0 43.4



#16
Acetone
Concen: 250.057 ug/l
RT: 2.374 min Scan# 211
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 43 Resp: 529575
Ion Ratio Lower Upper
43 100
58 30.7 22.6 33.8









#17
Carbon Disulfide
Concen: 43.385 ug/l
RT: 2.514 min Scan# 211
Delta R.T. 0.006 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 411279
Ion Ratio Lower Upper
76 100
78 9.2 7.0 10.6



#18
Methyl Acetate
Concen: 48.592 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.006 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 43 Resp: 318428
Ion Ratio Lower Upper
43 100
74 23.0 18.2 27.4









#19
Methyl tert-butyl Ether
Concen: 50.356 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 73 Resp: 586002
Ion Ratio Lower Upper
73 100
57 22.1 17.4 26.0



#20
Methylene Chloride
Concen: 47.647 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 84 Resp: 181816
Ion Ratio Lower Upper
84 100
49 128.7 101.1 151.7
51 39.2 31.0 46.6
86 64.5 52.6 79.0









#21
trans-1,2-Dichloroethene
Concen: 48.669 ug/l
RT: 3.093 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 96 Resp: 168152
Ion Ratio Lower Upper
96 100
61 136.2 106.7 160.1
98 62.5 50.0 75.0



#22
Diisopropyl ether
Concen: 52.809 ug/l
RT: 3.758 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 45 Resp: 647527
Ion Ratio Lower Upper
45 100
43 74.1 64.6 97.0
87 26.4 21.9 32.9
59 11.0 9.2 13.8









#23
 Vinyl Acetate
 Concen: 256.236 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. 0.006 min
 Lab File: VX034359.D
 Acq: 07 Mar 2023 09:44




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 43 Resp: 2517254
 Ion Ratio Lower Upper
 43 100
 86 10.6 8.4 12.6



#24
 1,1-Dichloroethane
 Concen: 49.232 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.006 min
 Lab File: VX034359.D
 Acq: 07 Mar 2023 09:44

Tgt Ion: 63 Resp: 331401
 Ion Ratio Lower Upper
 63 100
 98 7.0 3.8 11.2
 100 4.9 2.4 7.2





#25

2-Butanone

Concen: 242.515 ug/l

RT: 4.550 min Scan# 5

Delta R.T. 0.006 min

Lab File: VX034359.D

Acq: 07 Mar 2023 09:44

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 43 Resp: 743139

Ion Ratio Lower Upper

43 100

72 24.9 19.4 29.2



#26

2,2-Dichloropropane

Concen: 47.353 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.000 min

Lab File: VX034359.D




Acq: 07 Mar 2023 09:44


Tgt Ion: 77 Resp: 288754

Ion Ratio Lower Upper

77 100

97 22.8 11.9 35.5









#27
cis-1,2-Dichloroethene
Concen: 50.242 ug/l
RT: 4.483 min Scan# 511
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 96 Resp: 205558
Ion Ratio Lower Upper
96 100
61 148.2 0.0 277.6
98 64.7 0.0 128.0



#28
Bromochloromethane
Concen: 50.579 ug/l
RT: 4.891 min Scan# 624
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 49 Resp: 151883
Ion Ratio Lower Upper
49 100
129 1.9 0.0 4.6
130 73.5 68.5 102.7





#29

Tetrahydrofuran

Concen: 239.416 ug/l

RT: 5.001 min Scan# 641

Delta R.T. 0.006 min

Lab File: VX034359.D

Acq: 07 Mar 2023 09:44

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 42 Resp: 447554

Ion Ratio Lower Upper

42 100

72 42.7 34.8 52.2

71 39.8 32.2 48.2



#30

Chloroform

Concen: 52.079 ug/l

RT: 5.087 min Scan# 656

Delta R.T. -0.006 min

Lab File: VX034359.D




Acq: 07 Mar 2023 09:44


Tgt Ion: 83 Resp: 336627

Ion Ratio Lower Upper

83 100

85 66.4 52.1 78.1









#31
Cyclohexane
Concen: 48.643 ug/l
RT: 5.471 min Scan# 719
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 298161
Ion Ratio Lower Upper
56 100
69 28.8 24.6 37.0
84 82.7 71.0 106.6



#32
1,1,1-Trichloroethane
Concen: 50.219 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 97 Resp: 290985
Ion Ratio Lower Upper
97 100
99 64.5 52.3 78.5
61 45.2 35.2 52.8









#33
1,2-Dichloroethane-d4
Concen: 50.929 ug/l
RT: 5.952 min Scan# 798
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 65 Resp: 209091
Ion Ratio Lower Upper
65 100
67 52.4 0.0 104.8



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 114 Resp: 500792
Ion Ratio Lower Upper
114 100
63 21.4 0.0 37.4
88 16.3 0.0 30.6





#35

Dibromofluoromethane

Concen: 51.641 ug/l

RT: 5.379 min Scan# 705

Delta R.T. -0.006 min

Lab File: VX034359.D

Acq: 07 Mar 2023 09:44

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:113 Resp: 168384

Ion Ratio Lower Upper

113 100

111 102.4 83.0 124.6

192 19.0 19.0 28.4



#36

1,1-Dichloropropene

Concen: 51.001 ug/l

RT: 5.690 min Scan# 755

Delta R.T. 0.000 min

Lab File: VX034359.D

Acq: 07 Mar 2023 09:44




Tgt Ion: 75 Resp: 245934


Ion Ratio Lower Upper

75 100

110 35.6 18.8 56.4

77 30.6 25.1 37.7









#37
Ethyl Acetate
Concen: 50.420 ug/l
RT: 4.709 min Scan# 594
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 271724
Ion Ratio Lower Upper
43 100
61 13.5 11.1 16.7
70 11.0 8.5 12.7



#38
Carbon Tetrachloride
Concen: 50.848 ug/l
RT: 5.678 min Scan# 753
Delta R.T. 0.006 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 117 Resp: 252315
Ion Ratio Lower Upper
117 100
119 93.6 78.0 117.0
121 30.7 24.4 36.6





#39

Methylcyclohexane

Concen: 49.006 ug/l

RT: 7.379 min Scan# 1032

Delta R.T. 0.000 min

Lab File: VX034359.D

Acq: 07 Mar 2023 09:44

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 83 Resp: 304112

Ion Ratio Lower Upper

83 100

55 77.4 61.8 92.6

98 49.6 39.3 58.9



#40

Benzene

Concen: 50.998 ug/l

RT: 6.032 min Scan# 811

Delta R.T. -0.006 min

Lab File: VX034359.D




Acq: 07 Mar 2023 09:44


Tgt Ion: 78 Resp: 728782

Ion Ratio Lower Upper

78 100

77 23.8 19.0 28.4









#41
Methacrylonitrile
Concen: 51.135 ug/l
RT: 4.916 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 41 Resp: 153540
Ion Ratio Lower Upper
41 100
39 54.9 43.8 65.6
67 69.6 54.8 82.2
52 29.1 23.0 34.6



#42
1,2-Dichloroethane
Concen: 55.851 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 62 Resp: 280709
Ion Ratio Lower Upper
62 100
98 8.9 0.0 20.2









#43
Isopropyl Acetate
Concen: 50.982 ug/l
RT: 6.336 min Scan# 861
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 484851
Ion Ratio Lower Upper
43 100
61 19.7 15.7 23.5
87 12.5 10.6 15.8



#44
Trichloroethene
Concen: 51.267 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 130 Resp: 188005
Ion Ratio Lower Upper
130 100
95 97.4 0.0 182.6





#45

1,2-Dichloropropane

Concen: 51.721 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX034359.D

Acq: 07 Mar 2023 09:44

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 63 Resp: 196758

Ion Ratio Lower Upper

63 100

65 31.3 24.2 36.2



#46

Dibromomethane

Concen: 53.222 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.000 min

Lab File: VX034359.D

Acq: 07 Mar 2023 09:44




Tgt Ion: 93 Resp: 133248


Ion Ratio Lower Upper

93 100

95 82.3 67.6 101.4

174 96.6 97.4 146.0#





#47

Bromodichloromethane

Concen: 52.299 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. 0.000 min

Lab File: VX034359.D

Acq: 07 Mar 2023 09:44

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 83 Resp: 268711

Ion Ratio Lower Upper

83 100

85 64.2 49.8 74.6

127 8.4 7.0 10.4



#48

Methyl methacrylate

Concen: 51.247 ug/l

RT: 7.690 min Scan# 1083

Delta R.T. 0.000 min

Lab File: VX034359.D

Acq: 07 Mar 2023 09:44




Tgt Ion: 41 Resp: 231323


Ion Ratio Lower Upper

41 100

69 77.6 63.1 94.7

39 59.2 43.2 64.8





#49

1,4-Dioxane

Concen: 991.449 ug/l

RT: 7.653 min Scan# 1077

Delta R.T. 0.000 min

Lab File: VX034359.D

Acq: 07 Mar 2023 09:44

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 88 Resp: 89264

Ion Ratio Lower Upper

88 100

43 34.4 0.0 0.0#

58 72.4 56.5 84.7



#50

Toluene-d8

Concen: 51.221 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX034359.D




Acq: 07 Mar 2023 09:44


Tgt Ion: 98 Resp: 599631

Ion Ratio Lower Upper

98 100

100 65.7 53.6 80.4





#51

4-Methyl-2-Pentanone

Concen: 268.855 ug/l

RT: 8.568 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX034359.D

Acq: 07 Mar 2023 09:44

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 43 Resp: 1406731

Ion Ratio Lower Upper

43 100

58 38.2 31.2 46.8



#52

Toluene

Concen: 53.678 ug/l

RT: 8.714 min Scan# 1251

Delta R.T. 0.000 min

Lab File: VX034359.D




Acq: 07 Mar 2023 09:44


Tgt Ion: 92 Resp: 464791

Ion Ratio Lower Upper

92 100

91 171.5 135.5 203.3





#53

t-1,3-Dichloropropene

Concen: 53.403 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX034359.D

Acq: 07 Mar 2023 09:44

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 75 Resp: 315834

Ion Ratio Lower Upper

75 100

77 31.3 25.5 38.3



#54

cis-1,3-Dichloropropene

Concen: 51.545 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX034359.D

Acq: 07 Mar 2023 09:44




Tgt Ion: 75 Resp: 326355


Ion Ratio Lower Upper

75 100

77 32.1 25.0 37.6

39 53.5 38.6 58.0









#55
1,1,2-Trichloroethane
Concen: 55.501 ug/l
RT: 9.147 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 97 Resp: 193982




Ion	Ratio	Lower	Upper
97	100		
83	86.6	67.4	101.0
85	54.5	43.2	64.8
99	60.3	50.2	75.2




#56
Ethyl methacrylate
Concen: 54.652 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 69 Resp: 323134

Ion	Ratio	Lower	Upper
69	100		
41	70.5	55.1	82.7
39	40.9	28.5	42.7









#57
1,3-Dichloropropane
Concen: 55.174 ug/l
RT: 9.305 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 335508
Ion Ratio Lower Upper
76 100
78 32.1 25.6 38.4



#58
2-Chloroethyl Vinyl ether
Concen: 260.668 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 63 Resp: 667640
Ion Ratio Lower Upper
63 100
106 28.1 24.5 36.7





#59

2-Hexanone

Concen: 276.567 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX034359.D

Acq: 07 Mar 2023 09:44

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 43 Resp: 1102831

Ion Ratio Lower Upper

43 100

58 52.7 27.4 82.0



#60

Dibromochloromethane

Concen: 54.769 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX034359.D




Acq: 07 Mar 2023 09:44


Tgt Ion: 129 Resp: 211511

Ion Ratio Lower Upper

129 100

127 77.5 38.8 116.3









#61
1,2-Dibromoethane
Concen: 55.705 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:107 Resp: 205907
Ion Ratio Lower Upper
107 100
109 93.2 74.2 111.2



#62
4-Bromofluorobenzene
Concen: 56.477 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 95 Resp: 252899
Ion Ratio Lower Upper
95 100
174 73.2 0.0 184.4
176 70.9 0.0 178.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:117 Resp: 463756
Ion Ratio Lower Upper
117 100
82 55.1 40.6 60.8
119 32.9 25.6 38.4



#64
Tetrachloroethene
Concen: 49.148 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.006 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion:164 Resp: 160866
Ion Ratio Lower Upper
164 100
166 126.5 100.2 150.2
129 96.1 69.2 103.8
131 93.8 67.2 100.8









#65
Chlorobenzene
Concen: 51.611 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:112 Resp: 498920
Ion Ratio Lower Upper
112 100
114 32.2 26.2 39.4



#66
1,1,1,2-Tetrachloroethane
Concen: 51.829 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.006 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion:131 Resp: 193961
Ion Ratio Lower Upper
131 100
133 94.5 48.4 145.0
119 64.2 31.9 95.7









#67
Ethyl Benzene
Concen: 51.983 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 902797
Ion Ratio Lower Upper
91 100
106 31.7 25.8 38.8



#68
m/p-Xylenes
Concen: 103.170 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 106 Resp: 685691
Ion Ratio Lower Upper
106 100
91 206.5 156.4 234.6









#69
o-Xylene
Concen: 51.991 ug/l
RT: 10.640 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:106 Resp: 344004
Ion Ratio Lower Upper
106 100
91 216.2 103.1 309.3



#70
Styrene
Concen: 53.357 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. -0.006 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion:104 Resp: 574183
Ion Ratio Lower Upper
104 100
78 53.3 37.5 56.3
103 54.4 43.3 64.9









#71
Bromoform
Concen: 51.264 ug/l
RT: 10.799 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:173 Resp: 156595
Ion Ratio Lower Upper
173 100
175 47.8 24.1 72.3
254 0.1 0.1 0.1



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion:152 Resp: 235275
Ion Ratio Lower Upper
152 100
115 86.9 38.7 116.1
150 174.0 0.0 348.2









#73
Isopropylbenzene
Concen: 50.071 ug/l
RT: 10.964 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:105 Resp: 897822
Ion Ratio Lower Upper
105 100
120 26.7 13.6 40.8



#74
N-ethyl acetate
Concen: 49.103 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 43 Resp: 442338
Ion Ratio Lower Upper
43 100
70 40.7 34.8 52.2
55 22.7 19.4 29.0
61 23.6 19.4 29.2









#75
1,1,2,2-Tetrachloroethane
Concen: 48.775 ug/l
RT: 11.214 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 299132
Ion Ratio Lower Upper
83 100
131 10.1 5.5 16.7
85 64.3 32.3 96.9



#76
1,2,3-Trichloropropane
Concen: 58.733 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 75 Resp: 347709
Ion Ratio Lower Upper
75 100
77 32.3 21.8 65.4









#77
Bromobenzene
Concen: 50.254 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:156 Resp: 216851
Ion Ratio Lower Upper
156 100
77 171.6 68.1 204.3
158 96.0 48.0 144.2



#78
n-propylbenzene
Concen: 50.819 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 91 Resp: 1059776
Ion Ratio Lower Upper
91 100
120 23.3 12.3 36.8





#79

2-Chlorotoluene

Concen: 50.053 ug/l

RT: 11.366 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX034359.D

Acq: 07 Mar 2023 09:44

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 91 Resp: 640039

Ion Ratio Lower Upper

91 100

126 33.3 18.1 54.1



#80

1,3,5-Trimethylbenzene

Concen: 50.018 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX034359.D




Acq: 07 Mar 2023 09:44


Tgt Ion: 105 Resp: 754042

Ion Ratio Lower Upper

105 100

120 48.4 25.6 76.8









#81
trans-1,4-Dichloro-2-butene
Concen: 44.923 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 75 Resp: 103528
Ion Ratio Lower Upper
75 100
53 99.0 83.0 124.6
89 44.9 37.7 56.5



#82
4-Chlorotoluene
Concen: 50.630 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.006 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 91 Resp: 737398
Ion Ratio Lower Upper
91 100
126 28.6 15.8 47.4









#83
tert-Butylbenzene
Concen: 48.789 ug/l
RT: 11.713 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:119 Resp: 753300
Ion Ratio Lower Upper
119 100
91 61.6 26.9 80.7
134 23.5 12.3 36.8



#84
1,2,4-Trimethylbenzene
Concen: 50.317 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion:105 Resp: 766973
Ion Ratio Lower Upper
105 100
120 44.3 23.4 70.2









#85
 sec-Butylbenzene
 Concen: 50.322 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX034359.D
 Acq: 07 Mar 2023 09:44




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:105 Resp: 943543
 Ion Ratio Lower Upper
 105 100
 134 20.2 10.8 32.3



#86
 p-Isopropyltoluene
 Concen: 50.328 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.006 min
 Lab File: VX034359.D
 Acq: 07 Mar 2023 09:44

Tgt Ion:119 Resp: 780954
 Ion Ratio Lower Upper
 119 100
 134 26.7 13.6 40.8
 91 25.9 11.0 33.0





#87

1,3-Dichlorobenzene

Concen: 50.615 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX034359.D

Acq: 07 Mar 2023 09:44

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:146 Resp: 409605

Ion Ratio Lower Upper

146 100

111 41.6 18.6 55.6

148 64.0 31.9 95.5



#88

1,4-Dichlorobenzene

Concen: 50.121 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. 0.000 min

Lab File: VX034359.D

Acq: 07 Mar 2023 09:44




Tgt Ion:146 Resp: 409115


Ion Ratio Lower Upper

146 100

111 41.1 17.8 53.4

148 64.2 32.1 96.3









#89
n-Butylbenzene
Concen: 50.992 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 704743
Ion Ratio Lower Upper
91 100
92 55.0 27.5 82.5
134 25.9 14.3 42.9



#90
Hexachloroethane
Concen: 47.121 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 117 Resp: 152880
Ion Ratio Lower Upper
117 100
201 69.6 45.1 135.3









#91
1,2-Dichlorobenzene
Concen: 50.259 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:146 Resp: 397943
Ion Ratio Lower Upper
146 100
111 43.3 19.3 57.8
148 63.9 32.3 96.9



#92
1,2-Dibromo-3-Chloropropane
Concen: 46.674 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion: 75 Resp: 68259
Ion Ratio Lower Upper
75 100
155 77.9 51.7 155.1
157 101.1 67.5 202.3









#93
1,2,4-Trichlorobenzene
Concen: 49.042 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.006 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:180 Resp: 249297
Ion Ratio Lower Upper
180 100
182 96.7 47.9 143.6
145 33.0 14.1 42.2



#94
Hexachlorobutadiene
Concen: 48.527 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion:225 Resp: 102262
Ion Ratio Lower Upper
225 100
223 63.2 31.3 93.9
227 63.1 32.3 96.9









#95
Naphthalene
Concen: 49.526 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:128 Resp: 848109
Ion Ratio Lower Upper
128 100
127 12.9 10.3 15.5
129 10.9 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 49.788 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX034359.D
Acq: 07 Mar 2023 09:44

Tgt Ion:180 Resp: 251264
Ion Ratio Lower Upper
180 100
182 95.2 47.6 142.9
145 34.3 15.1 45.3

