

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030723\
 Data File : VX034365.D
 Acq On : 07 Mar 2023 12:14
 Operator : JC/MD
 Sample : 01821-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-IDW-S-1

Quant Time: Mar 08 01:01:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	250710	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	428573	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	386301	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	168537	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	160156	47.433	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	94.860%	
35) Dibromofluoromethane	5.385	113	134499	48.200	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	96.400%	
50) Toluene-d8	8.647	98	515140	51.418	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	102.840%	
62) 4-Bromofluorobenzene	11.079	95	180308	47.051	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	94.100%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.953	59	6967	9.225	ug/l #	92
16) Acetone	2.380	43	44519	25.560	ug/l	92
20) Methylene Chloride	2.794	84	11769	3.750	ug/l	93
25) 2-Butanone	4.568	43	24251	9.623	ug/l	98
37) Ethyl Acetate	4.721	43	65864	14.281	ug/l	99
43) Isopropyl Acetate	6.342	43	271821	33.398	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030723\
Data File : VX034365.D
Acq On : 07 Mar 2023 12:14
Operator : JC/MD
Sample : 01821-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-IDW-S-1

Quant Time: Mar 08 01:01:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

