

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030723\
 Data File : VX034366.D
 Acq On : 07 Mar 2023 12:37
 Operator : JC/MD
 Sample : 01825-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HH-TP1-0306

Quant Time: Mar 08 01:01:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	256871	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	435037	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	389655	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	163922	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	162611	47.005	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	94.000%	
35) Dibromofluoromethane	5.385	113	136559	48.211	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	96.420%	
50) Toluene-d8	8.647	98	531453	52.259	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	104.520%	
62) 4-Bromofluorobenzene	11.079	95	181261	46.597	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	93.200%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.947	59	7051	9.113	ug/l #	95
16) Acetone	2.380	43	34569	19.371	ug/l	90
20) Methylene Chloride	2.788	84	12151	3.779	ug/l	97
25) 2-Butanone	4.562	43	20953	8.115	ug/l	96
37) Ethyl Acetate	4.721	43	62715	13.396	ug/l	98
43) Isopropyl Acetate	6.342	43	260707	31.557	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030723\
Data File : VX034366.D
Acq On : 07 Mar 2023 12:37
Operator : JC/MD
Sample : 01825-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HH-TP1-0306

Quant Time: Mar 08 01:01:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

