

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030723\
 Data File : VX034367.D
 Acq On : 07 Mar 2023 13:01
 Operator : JC/MD
 Sample : 01825-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HH-TP2-0306

Quant Time: Mar 08 01:01:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	258791	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	440248	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	401025	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	173689	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	165054	47.357	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	94.720%	
35) Dibromofluoromethane	5.391	113	139043	48.507	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	97.020%	
50) Toluene-d8	8.647	98	535958	52.078	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	104.160%	
62) 4-Bromofluorobenzene	11.079	95	188089	47.780	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	95.560%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.953	59	5779	7.413	ug/l #	86
16) Acetone	2.380	43	35567	19.783	ug/l #	89
20) Methylene Chloride	2.788	84	12246	3.780	ug/l	94
25) 2-Butanone	4.568	43	19867	7.637	ug/l	96
37) Ethyl Acetate	4.715	43	61001	12.876	ug/l	99
43) Isopropyl Acetate	6.342	43	261966	31.334	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030723\
Data File : VX034367.D
Acq On : 07 Mar 2023 13:01
Operator : JC/MD
Sample : 01825-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HH-TP2-0306

Quant Time: Mar 08 01:01:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

