

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030723\
 Data File : VX034370.D
 Acq On : 07 Mar 2023 14:56
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 08 01:02:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	264724	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	447225	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	392971	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	168740	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	165840	46.516	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.040%
35) Dibromofluoromethane	5.379	113	141081	48.450	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.900%
50) Toluene-d8	8.647	98	541304	51.777	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.560%
62) 4-Bromofluorobenzene	11.079	95	181797	45.461	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030723\
Data File : VX034370.D
Acq On : 07 Mar 2023 14:56
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 08 01:02:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

