

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030723\
 Data File : VX034371.D
 Acq On : 07 Mar 2023 17:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 08 01:02:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	59	0.00
2 T	Dichlorodifluoromethane	0.432	0.443	-2.5	57	0.00
3 P	Chloromethane	0.622	0.619	0.5	57	0.00
4 C	Vinyl Chloride	0.566	0.631	-11.5#	63	0.00
5 T	Bromomethane	0.225	0.320	-42.2#	77	0.00
6 T	Chloroethane	0.285	0.360	-26.3#	69	0.00
7 T	Trichlorofluoromethane	0.766	0.974	-27.2#	73	0.00
8 T	Diethyl Ether	0.309	0.350	-13.3	69	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.539	0.549	-1.9	58	0.00
10 T	Methyl Iodide	0.657	0.666	-1.4	56	0.00
11 T	Tert butyl alcohol	0.151	0.133	11.9	53	0.00
12 CM	1,1-Dichloroethene	0.541	0.524	3.1#	56	0.00
13 T	Acrolein	0.069	0.074	-7.2	67	0.00
14 T	Allyl chloride	1.147	1.042	9.2	52	0.00
15 T	Acrylonitrile	0.329	0.323	1.8	56	0.00
16 T	Acetone	0.347	0.302	13.0	52	0.00
17 T	Carbon Disulfide	1.555	1.377	11.4	51	0.00
18 T	Methyl Acetate	1.075	1.078	-0.3	56	0.00
19 T	Methyl tert-butyl Ether	1.909	1.906	0.2	57	0.00
20 T	Methylene Chloride	0.626	0.596	4.8	57	0.00
21 T	trans-1,2-Dichloroethene	0.567	0.575	-1.4	57	0.00
22 T	Diisopropyl ether	2.011	2.137	-6.3	60	0.00
23 T	Vinyl Acetate	1.611	1.662	-3.2	59	0.00
24 P	1,1-Dichloroethane	1.104	1.108	-0.4	58	0.00
25 T	2-Butanone	0.503	0.477	5.2	55	0.00
26 T	2,2-Dichloropropane	1.000	0.976	2.4	57	0.00
27 T	cis-1,2-Dichloroethene	0.671	0.687	-2.4	58	0.00
28 T	Bromochloromethane	0.493	0.456	7.5	55	0.00
29 T	Tetrahydrofuran	0.307	0.304	1.0	57	0.00
30 C	Chloroform	1.060	1.117	-5.4#	60	0.00
31 T	Cyclohexane	1.005	1.015	-1.0	57	0.00
32 T	1,1,1-Trichloroethane	0.950	0.978	-2.9	59	0.00
33 S	1,2-Dichloroethane-d4	0.673	0.625	7.1	57	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	59	0.00
35 S	Dibromofluoromethane	0.326	0.312	4.3	58	0.00
36 T	1,1-Dichloropropene	0.481	0.503	-4.6	59	0.00
37 T	Ethyl Acetate	0.538	0.544	-1.1	56	0.00
38 T	Carbon Tetrachloride	0.495	0.508	-2.6	59	0.00
39 T	Methylcyclohexane	0.620	0.622	-0.3	56	0.00
40 TM	Benzene	1.427	1.479	-3.6	60	0.00
41 T	Methacrylonitrile	0.300	0.313	-4.3	58	0.00
42 TM	1,2-Dichloroethane	0.502	0.536	-6.8	61	0.00
43 T	Isopropyl Acetate	0.950	0.953	-0.3	57	0.00
44 TM	Trichloroethene	0.366	0.386	-5.5	60	0.00
45 C	1,2-Dichloropropane	0.380	0.394	-3.7#	59	0.00
46 T	Dibromomethane	0.250	0.261	-4.4	60	0.00
47 T	Bromodichloromethane	0.513	0.531	-3.5	59	0.00
48 T	Methyl methacrylate	0.451	0.458	-1.6	58	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030723\
 Data File : VX034371.D
 Acq On : 07 Mar 2023 17:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 08 01:02:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	61	0.00
50 S	Toluene-d8	1.169	1.145	2.1	59	0.00
51 T	4-Methyl-2-Pentanone	0.522	0.571	-9.4	62	0.00
52 CM	Toluene	0.865	0.952	-10.1#	63	0.00
53 T	t-1,3-Dichloropropene	0.590	0.620	-5.1	59	0.00
54 T	cis-1,3-Dichloropropene	0.632	0.644	-1.9	57	0.00
55 T	1,1,2-Trichloroethane	0.349	0.384	-10.0	63	0.00
56 T	Ethyl methacrylate	0.590	0.631	-6.9	60	0.00
57 T	1,3-Dichloropropane	0.607	0.661	-8.9	62	0.00
58 T	2-Chloroethyl Vinyl ether	0.256	0.270	-5.5	61	0.00
59 T	2-Hexanone	0.398	0.432	-8.5	62	0.00
60 T	Dibromochloromethane	0.386	0.418	-8.3	62	0.00
61 T	1,2-Dibromoethane	0.369	0.401	-8.7	63	0.00
62 S	4-Bromofluorobenzene	0.447	0.479	-7.2	64	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	62	0.00
64 T	Tetrachloroethene	0.353	0.358	-1.4	62	0.00
65 PM	Chlorobenzene	1.042	1.117	-7.2	65	0.00
66 T	1,1,1,2-Tetrachloroethane	0.403	0.419	-4.0	63	0.00
67 C	Ethyl Benzene	1.872	2.045	-9.2#	65	0.00
68 T	m/p-Xylenes	0.717	0.781	-8.9	66	0.00
69 T	o-Xylene	0.713	0.776	-8.8	66	0.00
70 T	Styrene	1.160	1.294	-11.6	67	0.00
71 P	Bromoform	0.329	0.334	-1.5	61	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	68	0.00
73 T	Isopropylbenzene	3.811	3.874	-1.7	68	0.00
74 T	N-amyl acetate	1.914	1.898	0.8	68	0.00
75 P	1,1,2,2-Tetrachloroethane	1.303	1.277	2.0	67	0.00
76 T	1,2,3-Trichloropropane	1.258	1.072	14.8	58	0.00
77 T	Bromobenzene	0.917	0.916	0.1	68	0.00
78 T	n-propylbenzene	4.432	4.623	-4.3	68	0.00
79 T	2-Chlorotoluene	2.718	2.756	-1.4	68	0.00
80 T	1,3,5-Trimethylbenzene	3.204	3.274	-2.2	68	0.00
81 T	trans-1,4-Dichloro-2-butene	0.490	0.438	10.6	60	0.00
82 T	4-Chlorotoluene	3.095	3.208	-3.7	70	0.00
83 T	tert-Butylbenzene	3.281	3.253	0.9	66	0.00
84 T	1,2,4-Trimethylbenzene	3.239	3.314	-2.3	69	0.00
85 T	sec-Butylbenzene	3.985	4.117	-3.3	68	0.00
86 T	p-Isopropyltoluene	3.298	3.417	-3.6	68	0.00
87 T	1,3-Dichlorobenzene	1.720	1.783	-3.7	69	0.00
88 T	1,4-Dichlorobenzene	1.735	1.799	-3.7	70	0.00
89 T	n-Butylbenzene	2.937	3.095	-5.4	68	0.00
90 T	Hexachloroethane	0.689	0.655	4.9	63	0.00
91 T	1,2-Dichlorobenzene	1.683	1.711	-1.7	68	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.311	0.284	8.7	61	0.00
93 T	1,2,4-Trichlorobenzene	1.080	1.089	-0.8	65	0.00
94 T	Hexachlorobutadiene	0.448	0.437	2.5	64	0.00
95 T	Naphthalene	3.639	3.607	0.9	66	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030723\
Data File : VX034371.D
Acq On : 07 Mar 2023 17:05
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 08 01:02:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.073	1.068	0.5	66	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6