



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030723\
Data File : VX034358.D
Acq On : 07 Mar 2023 08:27
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Title : SW846 8260
Last Update : Fri Feb 17 12:33:35 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	8395	PASS
75	95	30	60	52.7	22963	PASS
95	95	100	100	100.0	43611	PASS
96	95	5	9	6.3	2760	PASS
173	174	0.00	2	0.5	169	PASS
174	95	50	100	76.4	33336	PASS
175	174	5	9	7.3	2436	PASS
176	174	95	101	97.2	32413	PASS
177	176	5	9	6.9	2247	PASS