

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030819\
 Data File : VX007932.D
 Acq On : 08 Mar 2019 15:36
 Operator : JC/SP
 Sample : K1810-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 55-ST-3

Quant Time: Mar 09 04:37:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X030819W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Mar 08 12:01:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.03	128	43981	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.86	114	228452	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	197454	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.07	65	87624	30.31	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	101.03%
60) 4-Bromofluorobenzene	11.13	95	88838	28.19	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	93.97%
63) Toluene-d8	8.71	98	268040	29.66	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	98.87%

Target Compounds

					Ovalue
15) Acetone	2.45	58	1217	1.967	ug/l 88
25) Chloroform	5.21	83	6145	1.251	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030819\
Data File : VX007932.D
Acq On : 08 Mar 2019 15:36
Operator : JC/SP
Sample : K1810-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
55-ST-3

Quant Time: Mar 09 04:37:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X030819W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Mar 08 12:01:36 2019
Response via : Initial Calibration

