

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030819\
 Data File : VX007940.D
 Acq On : 08 Mar 2019 19:35
 Operator : JC/SP
 Sample : K1815-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 09 05:24:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	277742	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	433667	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	410994	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	196792	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	153561	52.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.82%	
35) Dibromofluoromethane	5.50	113	142116	51.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.46%	
50) Toluene-d8	8.71	98	526974	51.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.46%	
62) 4-Bromofluorobenzene	11.13	95	194078	54.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.68%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030819\
Data File : VX007940.D
Acq On : 08 Mar 2019 19:35
Operator : JC/SP
Sample : K1815-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 09 05:24:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

