

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030819\
 Data File : VX007941.D
 Acq On : 08 Mar 2019 20:02
 Operator : JC/SP
 Sample : K1815-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Mar 09 05:26:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	274670	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	429513	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	401020	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	195678	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	151622	52.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.66%	
35) Dibromofluoromethane	5.50	113	139294	50.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.40%	
50) Toluene-d8	8.71	98	519196	51.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.92%	
62) 4-Bromofluorobenzene	11.13	95	189906	53.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.36%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030819\
Data File : VX007941.D
Acq On : 08 Mar 2019 20:02
Operator : JC/SP
Sample : K1815-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Mar 09 05:26:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

