

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030819\
 Data File : VX007944.D
 Acq On : 08 Mar 2019 21:22
 Operator : JC/SP
 Sample : K1793-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE126D3-20190306

Quant Time: Mar 09 05:35:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	269820	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	425988	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	398609	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	190324	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	150489	52.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.74%	
35) Dibromofluoromethane	5.50	113	137772	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
50) Toluene-d8	8.71	98	514622	51.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.84%	
62) 4-Bromofluorobenzene	11.13	95	187582	53.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.92%	
Target Compounds						
16) Acetone	2.45	43	4549	2.761	ug/l #	85
44) Trichloroethene	7.21	130	9366	2.960	ug/l	91
64) Tetrachloroethene	9.34	164	7132	2.159	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030819\
Data File : VX007944.D
Acq On : 08 Mar 2019 21:22
Operator : JC/SP
Sample : K1793-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE126D3-20190306

Quant Time: Mar 09 05:35:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

