

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030819\
 Data File : VX007948.D
 Acq On : 08 Mar 2019 23:08
 Operator : JC/SP
 Sample : PB117725TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB117725TB

Quant Time: Mar 09 05:47:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	277669	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	432760	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	410587	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	201079	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	150128	51.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.50%	
35) Dibromofluoromethane	5.50	113	138001	49.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.70%	
50) Toluene-d8	8.71	98	532276	52.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.72%	
62) 4-Bromofluorobenzene	11.13	95	196483	55.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.24%	
Target Compounds						
16) Acetone	2.45	43	6682	3.941	ug/l	98
20) Methylene Chloride	2.86	84	5644	2.092	ug/l	96
43) Isopropyl Acetate	6.46	43	15353	2.231	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030819\
Data File : VX007948.D
Acq On : 08 Mar 2019 23:08
Operator : JC/SP
Sample : PB117725TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB117725TB

Quant Time: Mar 09 05:47:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

