

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
 Data File : VX021051.D
 Acq On : 08 Mar 2021 11:49
 Operator : JC/MD
 Sample : VX0308WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0308WBL02

Quant Time: Mar 09 01:03:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	225154	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	415395	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	407686	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	178516	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	148498	48.83	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.66%
35) Dibromofluoromethane	5.464	113	129462	51.30	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.60%
50) Toluene-d8	8.694	98	528335	53.69	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.38%
62) 4-Bromofluorobenzene	11.118	95	190950	52.48	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.96%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
Data File : VX021051.D
Acq On : 08 Mar 2021 11:49
Operator : JC/MD
Sample : VX0308WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0308WBL02

Quant Time: Mar 09 01:03:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

