

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
 Data File : VX021053.D
 Acq On : 08 Mar 2021 12:30
 Operator : JC/MD
 Sample : M1594-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-030521

Quant Time: Mar 09 01:04:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	213693	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	398841	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	394959	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	186869	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	142449	49.35	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.70%
35) Dibromofluoromethane	5.470	113	125065	51.61	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.22%
50) Toluene-d8	8.694	98	503089	53.24	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.48%
62) 4-Bromofluorobenzene	11.124	95	195724	56.02	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	112.04%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
Data File : VX021053.D
Acq On : 08 Mar 2021 12:30
Operator : JC/MD
Sample : M1594-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-030521

Quant Time: Mar 09 01:04:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

