

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
 Data File : VX021061.D
 Acq On : 08 Mar 2021 15:20
 Operator : JC/MD
 Sample : M1596-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-117T-20210305

Quant Time: Mar 09 01:05:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	217483	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.835	114	402487	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	396147	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	183434	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	143785	48.95	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.90%
35) Dibromofluoromethane	5.464	113	125172	51.19	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.38%
50) Toluene-d8	8.700	98	509231	53.41	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.82%
62) 4-Bromofluorobenzene	11.118	95	194503	55.17	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	110.34%
Target Compounds						
					Qvalue	
16) Acetone	2.423	43	3690	3.05	ug/l	97
27) cis-1,2-Dichloroethene	4.564	96	30869	9.80	ug/l	87
44) Trichloroethene	7.188	130	16413	5.44	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
Data File : VX021061.D
Acq On : 08 Mar 2021 15:20
Operator : JC/MD
Sample : M1596-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-117T-20210305

Quant Time: Mar 09 01:05:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

