

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
 Data File : VX021064.D
 Acq On : 08 Mar 2021 16:22
 Operator : JC/MD
 Sample : M1596-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-121.5N-20210305

Quant Time: Mar 09 01:06:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	221449	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	401534	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	406396	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	166464	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	145351	48.59	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.18%
35) Dibromofluoromethane	5.470	113	128264	52.58	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.16%
50) Toluene-d8	8.694	98	506116	53.21	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.42%
62) 4-Bromofluorobenzene	11.118	95	184586	52.48	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.96%
Target Compounds						
					Qvalue	
16) Acetone	2.423	43	3399	2.76	ug/l	99
27) cis-1,2-Dichloroethene	4.564	96	3450	1.08	ug/l	96
44) Trichloroethene	7.188	130	2949	0.98	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
Data File : VX021064.D
Acq On : 08 Mar 2021 16:22
Operator : JC/MD
Sample : M1596-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-121.5N-20210305

Quant Time: Mar 09 01:06:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

