

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
 Data File : VX021065.D
 Acq On : 08 Mar 2021 16:44
 Operator : JC/MD
 Sample : M1596-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-121.5T-20210305

Quant Time: Mar 09 01:07:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	224195	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.829	114	409968	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	398455	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	166738	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	147119	48.58	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	97.16%		
35) Dibromofluoromethane	5.464	113	130316	52.32	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	104.64%		
50) Toluene-d8	8.694	98	516179	53.15	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	106.30%		
62) 4-Bromofluorobenzene	11.118	95	185715	51.71	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	103.42%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.187	85	921	0.38	ug/l	93
9) 1,1,2-Trichlorotrifluo...	2.364	101	599	0.25	ug/l	95
16) Acetone	2.417	43	3869	3.10	ug/l	91
21) trans-1,2-Dichloroethene	3.140	96	968	0.34	ug/l	89
27) cis-1,2-Dichloroethene	4.564	96	78043	24.03	ug/l	89
44) Trichloroethene	7.188	130	35522	11.57	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
Data File : VX021065.D
Acq On : 08 Mar 2021 16:44
Operator : JC/MD
Sample : M1596-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-121.5T-20210305

Quant Time: Mar 09 01:07:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

