

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
 Data File : VX021066.D
 Acq On : 08 Mar 2021 17:05
 Operator : JC/MD
 Sample : M1596-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-125.9N-20210305

Quant Time: Mar 09 01:07:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	214617	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	399044	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	397547	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	173042	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	142381	49.12	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.24%
35) Dibromofluoromethane	5.464	113	124933	51.53	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.06%
50) Toluene-d8	8.694	98	504062	53.32	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.64%
62) 4-Bromofluorobenzene	11.124	95	186412	53.33	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.66%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.358	96	1970	0.82	ug/l	91
16) Acetone	2.417	43	4645	3.89	ug/l	96
21) trans-1,2-Dichloroethene	3.146	96	1996	0.74	ug/l	83
27) cis-1,2-Dichloroethene	4.564	96	63935	20.57	ug/l	88
44) Trichloroethene	7.188	130	147720	49.42	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
Data File : VX021066.D
Acq On : 08 Mar 2021 17:05
Operator : JC/MD
Sample : M1596-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-125.9N-20210305

Quant Time: Mar 09 01:07:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

