

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
 Data File : VX021067.D
 Acq On : 08 Mar 2021 17:26
 Operator : JC/MD
 Sample : M1596-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-125.9T-20210305

Quant Time: Mar 09 01:07:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	209558	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.829	114	388938	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	391399	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	182023	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	140160	49.52	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.04%
35) Dibromofluoromethane	5.464	113	122989	52.05	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.10%
50) Toluene-d8	8.694	98	489910	53.17	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.34%
62) 4-Bromofluorobenzene	11.118	95	189926	55.75	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	111.50%
Target Compounds						
16) Acetone	2.417	43	2925	2.51	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
Data File : VX021067.D
Acq On : 08 Mar 2021 17:26
Operator : JC/MD
Sample : M1596-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-125.9T-20210305

Quant Time: Mar 09 01:07:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

