

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
 Data File : VX021068.D
 Acq On : 08 Mar 2021 17:49
 Operator : JC/MD
 Sample : M1596-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-318.9-20210305

Quant Time: Mar 09 01:08:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	206393	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	385383	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	378483	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	171128	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	139501	50.04	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.08%	
35) Dibromofluoromethane	5.464	113	121066	51.70	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.40%	
50) Toluene-d8	8.694	98	482647	52.87	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.74%	
62) 4-Bromofluorobenzene	11.117	95	184798	54.74	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 109.48%	
Target Compounds						
16) Acetone	2.416	43	1474	1.28	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
Data File : VX021068.D
Acq On : 08 Mar 2021 17:49
Operator : JC/MD
Sample : M1596-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-318.9-20210305

Quant Time: Mar 09 01:08:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

