

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
 Data File : VX021069.D
 Acq On : 08 Mar 2021 18:12
 Operator : JC/MD
 Sample : M1596-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-388.9-20210305

Quant Time: Mar 09 01:08:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	203218	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	391689	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	379475	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	164187	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	140601	51.22	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.44%	
35) Dibromofluoromethane	5.464	113	123633	51.95	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.90%	
50) Toluene-d8	8.694	98	494564	53.30	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.60%	
62) 4-Bromofluorobenzene	11.118	95	181111	52.79	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 105.58%	
Target Compounds						
16) Acetone	2.411	43	2039	1.80	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
Data File : VX021069.D
Acq On : 08 Mar 2021 18:12
Operator : JC/MD
Sample : M1596-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-388.9-20210305

Quant Time: Mar 09 01:08:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

