

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
 Data File : VX021070.D
 Acq On : 08 Mar 2021 18:35
 Operator : JC/MD
 Sample : M1596-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GR-OF-20210305

Quant Time: Mar 09 01:08:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	219900	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	401300	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	394179	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	169090	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	145245	48.90	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.80%
35) Dibromofluoromethane	5.464	113	126388	51.84	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.68%
50) Toluene-d8	8.694	98	501226	52.72	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.44%
62) 4-Bromofluorobenzene	11.118	95	190825	54.28	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.56%
Target Compounds						
					Qvalue	
16) Acetone	2.417	43	1647	1.35	ug/l	97
27) cis-1,2-Dichloroethene	4.564	96	6008	1.89	ug/l	91
44) Trichloroethene	7.188	130	3606	1.20	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
Data File : VX021070.D
Acq On : 08 Mar 2021 18:35
Operator : JC/MD
Sample : M1596-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GR-OF-20210305

Quant Time: Mar 09 01:08:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

