

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
 Data File : VX021083.D
 Acq On : 09 Mar 2021 00:21
 Operator : JC/MD
 Sample : PB135028ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB135028ZHE#01

Quant Time: Mar 09 04:23:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	223014	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	410169	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	407378	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	180294	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	146530	48.64	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.28%
35) Dibromofluoromethane	5.464	113	129261	51.87	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.74%
50) Toluene-d8	8.694	98	514825	52.98	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.96%
62) 4-Bromofluorobenzene	11.118	95	193093	53.74	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.48%
Target Compounds						
16) Acetone	2.417	43	47119	37.96	ug/l	99
20) Methylene Chloride	2.835	84	16233	5.45	ug/l	97
25) 2-Butanone	4.629	43	10473	5.55	ug/l	99
43) Isopropyl Acetate	6.411	43	20360	3.23	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030821\
Data File : VX021083.D
Acq On : 09 Mar 2021 00:21
Operator : JC/MD
Sample : PB135028ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB135028ZHE#01

Quant Time: Mar 09 04:23:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

