

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030822\
 Data File : VX027334.D
 Acq On : 08 Mar 2022 15:53
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 09 00:14:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	405943	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	676116	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	625562	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	293763	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	224948	47.518	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.040%
35) Dibromofluoromethane	5.391	113	213402	48.915	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.820%
50) Toluene-d8	8.652	98	801520	48.664	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.320%
62) 4-Bromofluorobenzene	11.085	95	301271	47.993	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.980%
Target Compounds						
17) Carbon Disulfide	2.513	76	15223	1.424	ug/l	Qvalue # 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030822\
Data File : VX027334.D
Acq On : 08 Mar 2022 15:53
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 09 00:14:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

