

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX030822\
 Data File : WX027335.D
 Acq On : 08 Mar 2022 16:16
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVX030822

Quant Time: Mar 09 00:15:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	50.000	46.936	6.1	97	0.00
3 P	Chloromethane	50.000	42.506	15.0	93	0.00
4 C	Vinyl Chloride	50.000	44.586	10.8#	94	0.00
5 T	Bromomethane	50.000	49.234	1.5	100	0.00
6 T	Chloroethane	50.000	42.550	14.9	93	0.00
7 T	Trichlorofluoromethane	50.000	44.469	11.1	96	0.00
8 T	Diethyl Ether	50.000	44.601	10.8	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	44.049	11.9	96	0.00
10 T	Methyl Iodide	50.000	47.206	5.6	95	0.00
11 T	Tert butyl alcohol	250.000	227.050	9.2	95	0.00
12 CM	1,1-Dichloroethene	50.000	45.212	9.6#	95	0.00
13 T	Acrolein	250.000	232.484	7.0	105	0.00
14 T	Allyl chloride	50.000	44.760	10.5	94	0.00
15 T	Acrylonitrile	250.000	222.052	11.2	93	0.00
16 T	Acetone	250.000	256.448	-2.6	103	0.00
17 T	Carbon Disulfide	50.000	42.501	15.0	93	0.00
18 T	Methyl Acetate	50.000	44.024	12.0	94	0.00
19 T	Methyl tert-butyl Ether	50.000	45.378	9.2	96	0.00
20 T	Methylene Chloride	50.000	43.090	13.8	93	0.00
21 T	trans-1,2-Dichloroethene	50.000	43.408	13.2	95	0.00
22 T	Diisopropyl ether	50.000	45.387	9.2	94	0.00
23 T	Vinyl Acetate	250.000	232.771	6.9	95	0.00
24 P	1,1-Dichloroethane	50.000	44.221	11.6	93	0.00
25 T	2-Butanone	250.000	227.676	8.9	96	0.00
26 T	2,2-Dichloropropane	50.000	45.811	8.4	96	0.00
27 T	cis-1,2-Dichloroethene	50.000	44.062	11.9	94	0.00
28 T	Bromochloromethane	50.000	40.751	18.5	87	0.00
29 T	Tetrahydrofuran	250.000	223.309	10.7	94	0.00
30 C	Chloroform	50.000	43.925	12.2#	93	0.00
31 T	Cyclohexane	50.000	45.522	9.0	94	0.00
32 T	1,1,1-Trichloroethane	50.000	45.033	9.9	94	0.00
33 S	1,2-Dichloroethane-d4	50.000	42.182	15.6	89	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	102	0.00
35 S	Dibromofluoromethane	50.000	42.613	14.8	88	0.00
36 T	1,1-Dichloropropene	50.000	45.371	9.3	95	0.00
37 T	Ethyl Acetate	50.000	45.142	9.7	94	0.00
38 T	Carbon Tetrachloride	50.000	45.242	9.5	94	0.00
39 T	Methylcyclohexane	50.000	45.880	8.2	96	0.00
40 TM	Benzene	50.000	44.231	11.5	94	0.00
41 T	Methacrylonitrile	50.000	46.038	7.9	94	0.00
42 TM	1,2-Dichloroethane	50.000	45.852	8.3	94	0.00
43 T	Isopropyl Acetate	50.000	45.877	8.2	95	0.00
44 TM	Trichloroethene	50.000	44.869	10.3	95	0.00
45 C	1,2-Dichloropropane	50.000	45.007	10.0#	93	0.00
46 T	Dibromomethane	50.000	45.259	9.5	94	0.00
47 T	Bromodichloromethane	50.000	45.801	8.4	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX030822\
 Data File : WX027335.D
 Acq On : 08 Mar 2022 16:16
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVX030822

Quant Time: Mar 09 00:15:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : Sw846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48 T	Methyl methacrylate	50.000	46.443	7.1	95	0.00
49 T	1,4-Dioxane	1000.000	888.586	11.1	95	0.00
50 S	Toluene-d8	50.000	42.537	14.9	88	0.00
51 T	4-Methyl-2-Pentanone	250.000	231.734	7.3	94	0.00
52 CM	Toluene	50.000	45.823	8.4#	94	0.00
53 T	t-1,3-Dichloropropene	50.000	47.688	4.6	95	0.00
54 T	cis-1,3-Dichloropropene	50.000	47.345	5.3	94	0.00
55 T	1,1,2-Trichloroethane	50.000	45.064	9.9	94	0.00
56 T	Ethyl methacrylate	50.000	47.400	5.2	96	0.00
57 T	1,3-Dichloropropane	50.000	45.057	9.9	94	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	237.436	5.0	96	0.00
59 T	2-Hexanone	250.000	236.695	5.3	95	0.00
60 T	Dibromochloromethane	50.000	46.372	7.3	94	0.00
61 T	1,2-Dibromoethane	50.000	46.532	6.9	95	0.00
62 S	4-Bromofluorobenzene	50.000	42.650	14.7	88	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	101	0.00
64 T	Tetrachloroethene	50.000	43.678	12.6	94	0.00
65 PM	Chlorobenzene	50.000	44.755	10.5	94	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	46.305	7.4	96	0.00
67 C	Ethyl Benzene	50.000	45.721	8.6#	94	0.00
68 T	m/p-Xylenes	100.000	92.259	7.7	95	0.00
69 T	o-Xylene	50.000	46.203	7.6	95	0.00
70 T	Styrene	50.000	46.398	7.2	94	0.00
71 P	Bromoform	50.000	46.094	7.8	92	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	101	0.00
73 T	Isopropylbenzene	50.000	45.909	8.2	95	0.00
74 T	N-amyl acetate	50.000	47.168	5.7	94	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	43.849	12.3	93	0.00
76 T	1,2,3-Trichloropropane	50.000	45.925	8.2	94	0.00
77 T	Bromobenzene	50.000	45.122	9.8	94	0.00
78 T	n-propylbenzene	50.000	46.696	6.6	94	0.00
79 T	2-Chlorotoluene	50.000	45.012	10.0	94	0.00
80 T	1,3,5-Trimethylbenzene	50.000	45.554	8.9	94	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.617	4.8	94	0.00
82 T	4-Chlorotoluene	50.000	45.529	8.9	93	0.00
83 T	tert-Butylbenzene	50.000	46.116	7.8	94	0.00
84 T	1,2,4-Trimethylbenzene	50.000	46.081	7.8	94	0.00
85 T	sec-Butylbenzene	50.000	46.572	6.9	95	0.00
86 T	p-Isopropyltoluene	50.000	46.751	6.5	94	0.00
87 T	1,3-Dichlorobenzene	50.000	44.691	10.6	93	0.00
88 T	1,4-Dichlorobenzene	50.000	44.564	10.9	95	0.00
89 T	n-Butylbenzene	50.000	47.379	5.2	94	0.00
90 T	Hexachloroethane	50.000	45.742	8.5	93	0.00
91 T	1,2-Dichlorobenzene	50.000	44.170	11.7	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	44.721	10.6	92	0.00
93 T	1,2,4-Trichlorobenzene	50.000	44.989	10.0	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030822\
Data File : VX027335.D
Acq On : 08 Mar 2022 16:16
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX030822

Quant Time: Mar 09 00:15:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : Sw846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
94 T	Hexachlorobutadiene	50.000	46.163	7.7	97	0.00
95 T	Naphthalene	50.000	46.387	7.2	92	0.00
96 T	1,2,3-Trichlorobenzene	50.000	44.737	10.5	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6