

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030822\
 Data File : VX027347.D
 Acq On : 08 Mar 2022 21:30
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 09 00:19:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : Sw846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	50.000	51.114	-2.2	98	0.00
3 P	Chloromethane	50.000	48.449	3.1	99	0.00
4 C	Vinyl Chloride	50.000	49.632	0.7#	98	0.00
5 T	Bromomethane	50.000	53.462	-6.9	100	0.00
6 T	Chloroethane	50.000	51.768	-3.5	103	0.00
7 T	Trichlorofluoromethane	50.000	48.834	2.3	98	0.00
8 T	Diethyl Ether	50.000	50.206	-0.4	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.515	3.0	98	0.00
10 T	Methyl Iodide	50.000	51.737	-3.5	97	0.00
11 T	Tert butyl alcohol	250.000	247.669	0.9	97	0.00
12 CM	1,1-Dichloroethene	50.000	50.464	-0.9#	98	0.00
13 T	Acrolein	250.000	267.746	-7.1	112	0.00
14 T	Allyl chloride	50.000	49.334	1.3	96	0.00
15 T	Acrylonitrile	250.000	254.463	-1.8	99	0.00
16 T	Acetone	250.000	267.811	-7.1	100	0.00
17 T	Carbon Disulfide	50.000	47.015	6.0	96	0.00
18 T	Methyl Acetate	50.000	49.583	0.8	98	0.00
19 T	Methyl tert-butyl Ether	50.000	50.141	-0.3	98	0.00
20 T	Methylene Chloride	50.000	48.492	3.0	98	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.926	2.1	99	0.00
22 T	Diisopropyl ether	50.000	50.888	-1.8	98	0.00
23 T	Vinyl Acetate	250.000	258.659	-3.5	98	0.00
24 P	1,1-Dichloroethane	50.000	49.967	0.1	98	0.00
25 T	2-Butanone	250.000	252.644	-1.1	100	0.00
26 T	2,2-Dichloropropane	50.000	45.250	9.5	88	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.487	1.0	98	0.00
28 T	Bromochloromethane	50.000	49.242	1.5	97	0.00
29 T	Tetrahydrofuran	250.000	252.957	-1.2	99	0.00
30 C	Chloroform	50.000	49.895	0.2#	98	0.00
31 T	Cyclohexane	50.000	50.690	-1.4	97	0.00
32 T	1,1,1-Trichloroethane	50.000	50.283	-0.6	97	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.205	1.6	97	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	96	0.00
35 S	Dibromofluoromethane	50.000	49.310	1.4	96	0.00
36 T	1,1-Dichloropropene	50.000	49.375	1.3	97	0.00
37 T	Ethyl Acetate	50.000	49.551	0.9	98	0.00
38 T	Carbon Tetrachloride	50.000	50.120	-0.2	98	0.00
39 T	Methylcyclohexane	50.000	48.919	2.2	96	0.00
40 TM	Benzene	50.000	49.313	1.4	99	0.00
41 T	Methacrylonitrile	50.000	51.071	-2.1	98	0.00
42 TM	1,2-Dichloroethane	50.000	50.666	-1.3	98	0.00
43 T	Isopropyl Acetate	50.000	49.917	0.2	97	0.00
44 TM	Trichloroethene	50.000	49.244	1.5	98	0.00
45 C	1,2-Dichloropropane	50.000	50.213	-0.4#	98	0.00
46 T	Dibromomethane	50.000	50.578	-1.2	99	0.00
47 T	Bromodichloromethane	50.000	50.970	-1.9	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030822\
 Data File : VX027347.D
 Acq On : 08 Mar 2022 21:30
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 09 00:19:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : Sw846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48 T	Methyl methacrylate	50.000	51.261	-2.5	99	0.00
49 T	1,4-Dioxane	1000.000	992.280	0.8	100	0.00
50 S	Toluene-d8	50.000	49.193	1.6	96	0.00
51 T	4-Methyl-2-Pentanone	250.000	255.312	-2.1	97	0.00
52 CM	Toluene	50.000	50.719	-1.4#	98	0.00
53 T	t-1,3-Dichloropropene	50.000	51.248	-2.5	96	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.018	-2.0	96	0.00
55 T	1,1,2-Trichloroethane	50.000	50.647	-1.3	100	0.00
56 T	Ethyl methacrylate	50.000	51.627	-3.3	98	0.00
57 T	1,3-Dichloropropane	50.000	50.227	-0.5	99	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	257.270	-2.9	98	0.00
59 T	2-Hexanone	250.000	252.452	-1.0	96	0.00
60 T	Dibromochloromethane	50.000	51.521	-3.0	98	0.00
61 T	1,2-Dibromoethane	50.000	50.949	-1.9	98	0.00
62 S	4-Bromofluorobenzene	50.000	47.431	5.1	92	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	95	0.00
64 T	Tetrachloroethene	50.000	48.057	3.9	97	0.00
65 PM	Chlorobenzene	50.000	49.471	1.1	98	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.309	-2.6	100	0.00
67 C	Ethyl Benzene	50.000	50.681	-1.4#	98	0.00
68 T	m/p-Xylenes	100.000	101.311	-1.3	98	0.00
69 T	o-Xylene	50.000	50.678	-1.4	98	0.00
70 T	Styrene	50.000	50.744	-1.5	96	0.00
71 P	Bromoform	50.000	50.641	-1.3	95	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	92	0.00
73 T	Isopropylbenzene	50.000	50.861	-1.7	97	0.00
74 T	N-amyl acetate	50.000	50.928	-1.9	93	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.668	2.7	95	0.00
76 T	1,2,3-Trichloropropane	50.000	50.451	-0.9	95	0.00
77 T	Bromobenzene	50.000	49.841	0.3	95	0.00
78 T	n-propylbenzene	50.000	51.250	-2.5	95	0.00
79 T	2-Chlorotoluene	50.000	49.619	0.8	95	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.313	-0.6	95	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	50.592	-1.2	92	0.00
82 T	4-Chlorotoluene	50.000	50.033	-0.1	94	0.00
83 T	tert-Butylbenzene	50.000	51.102	-2.2	96	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.523	-1.0	94	0.00
85 T	sec-Butylbenzene	50.000	50.920	-1.8	95	0.00
86 T	p-Isopropyltoluene	50.000	50.955	-1.9	94	0.00
87 T	1,3-Dichlorobenzene	50.000	49.303	1.4	95	0.00
88 T	1,4-Dichlorobenzene	50.000	48.759	2.5	96	0.00
89 T	n-Butylbenzene	50.000	50.025	-0.0	91	0.00
90 T	Hexachloroethane	50.000	51.200	-2.4	95	0.00
91 T	1,2-Dichlorobenzene	50.000	49.202	1.6	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	53.917	-7.8	101	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.373	1.3	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030822\
Data File : VX027347.D
Acq On : 08 Mar 2022 21:30
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 09 00:19:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : Sw846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
94 T	Hexachlorobutadiene	50.000	48.140	3.7	93	0.00
95 T	Naphthalene	50.000	51.892	-3.8	95	0.00
96 T	1,2,3-Trichlorobenzene	50.000	50.504	-1.0	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6