



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030822\
 Data File : VX027327.D
 Acq On : 08 Mar 2022 12:37
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Title : SW846 8260
 Last Update : Tue Mar 08 15:54:34 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.3	16961	PASS
75	95	30	60	50.0	52219	PASS
95	95	100	100	100.0	104360	PASS
96	95	5	9	6.9	7205	PASS
173	174	0.00	2	1.0	778	PASS
174	95	50	100	75.1	78424	PASS
175	174	5	9	7.4	5765	PASS
176	174	95	101	96.6	75741	PASS
177	176	5	9	6.7	5051	PASS