

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030823\
 Data File : VX034375.D
 Acq On : 08 Mar 2023 10:37
 Operator : JC/MD
 Sample : VX0308WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0308WBL01

Quant Time: Mar 09 03:56:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	267479	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	455791	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	397522	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	168167	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	169962	47.181	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.360%
35) Dibromofluoromethane	5.385	113	142196	47.916	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.840%
50) Toluene-d8	8.647	98	543461	51.006	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.020%
62) 4-Bromofluorobenzene	11.079	95	185926	45.620	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030823\
Data File : VX034375.D
Acq On : 08 Mar 2023 10:37
Operator : JC/MD
Sample : VX0308WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0308WBL01

Quant Time: Mar 09 03:56:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

