

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030823\
 Data File : VX034378.D
 Acq On : 08 Mar 2023 12:16
 Operator : JC/MD
 Sample : 01820-01
 Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HR-3-03032023

Quant Time: Mar 09 03:56:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	251572	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	424494	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	377411	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	183860	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	156478	46.185	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.360%
35) Dibromofluoromethane	5.385	113	127394	46.093	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.180%
50) Toluene-d8	8.647	98	513572	51.755	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.500%
62) 4-Bromofluorobenzene	11.085	95	186709	49.190	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030823\
Data File : VX034378.D
Acq On : 08 Mar 2023 12:16
Operator : JC/MD
Sample : 01820-01
Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HR-3-03032023

Quant Time: Mar 09 03:56:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

