

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030823\
 Data File : VX034386.D
 Acq On : 08 Mar 2023 16:19
 Operator : JC/MD
 Sample : 01843-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-20230306

Quant Time: Mar 09 04:00:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	255513	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	430456	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	378013	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	156378	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	161751	47.005	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.000%
35) Dibromofluoromethane	5.385	113	133017	47.461	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.920%
50) Toluene-d8	8.647	98	520262	51.703	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.400%
62) 4-Bromofluorobenzene	11.079	95	171808	44.637	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030823\
Data File : VX034386.D
Acq On : 08 Mar 2023 16:19
Operator : JC/MD
Sample : 01843-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-20230306

Quant Time: Mar 09 04:00:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

