

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030823\
 Data File : VX034388.D
 Acq On : 08 Mar 2023 17:06
 Operator : JC/MD
 Sample : 01843-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB01-20230308

Quant Time: Mar 09 04:01:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	256286	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	431298	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	376303	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	160343	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	163321	47.318	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.640%
35) Dibromofluoromethane	5.385	113	134721	47.975	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.940%
50) Toluene-d8	8.646	98	520401	51.615	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.240%
62) 4-Bromofluorobenzene	11.079	95	175301	45.456	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.920%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	9243	5.191	ug/l	92
30) Chloroform	5.098	83	1383	0.254	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030823\
Data File : VX034388.D
Acq On : 08 Mar 2023 17:06
Operator : JC/MD
Sample : 01843-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB01-20230308

Quant Time: Mar 09 04:01:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

