

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030823\
 Data File : VX034391.D
 Acq On : 08 Mar 2023 18:17
 Operator : JC/MD
 Sample : 01843-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D4-20230306

Quant Time: Mar 09 04:02:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 17 12:33:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	246808	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	416734	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	376725	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	162254	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	159209	47.898	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	95.800%	
35) Dibromofluoromethane	5.385	113	128940	47.521	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	95.040%	
50) Toluene-d8	8.647	98	504475	51.784	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	103.560%	
62) 4-Bromofluorobenzene	11.079	95	176006	47.233	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	94.460%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.167	85	1006	0.472	ug/l	75
7) Trichlorofluoromethane	1.892	101	1228	0.325	ug/l #	78
9) 1,1,2-Trichlorotrifluo...	2.331	101	13077	4.911	ug/l	92
12) 1,1-Dichloroethene	2.325	96	2341	0.877	ug/l #	81
27) cis-1,2-Dichloroethene	4.495	96	5581	1.685	ug/l	94
44) Trichloroethene	7.123	130	649390	212.799	ug/l	94
52) Toluene	8.720	92	8406	1.167	ug/l	97
64) Tetrachloroethene	9.275	164	16636	6.257	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030823\
Data File : VX034391.D
Acq On : 08 Mar 2023 18:17
Operator : JC/MD
Sample : 01843-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20230306

Quant Time: Mar 09 04:02:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X021723W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 17 12:33:35 2023
Response via : Initial Calibration

