

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030824\
 Data File : VX040614.D
 Acq On : 07 Mar 2024 09:48
 Operator : JC/MD
 Sample : VX0308MBL01
 Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK662

Quant Time: Mar 08 01:28:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML030724WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 08 01:27:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	153295	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.049	117	144951	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	71586	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	68986	51.338	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 102.680%	
7) Chloroethane-d5	1.666	69	51788	48.273	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 96.540%	
11) 1,1-Dichloroethene-d2	2.306	65	32795	53.016	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 106.040%	
21) 2-Butanone-d5	4.446	46	113660	110.225	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 110.220%	
24) Chloroform-d	5.050	84	133432	53.470	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 106.940%	
26) 1,2-Dichloroethane-d4	5.952	65	101896	54.691	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 109.380%	
32) Benzene-d6	5.970	84	251587	56.574	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 113.140%	
36) 1,2-Dichloropropane-d6	7.306	67	72222	52.401	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 104.800%	
41) Toluene-d8	8.647	98	222431	51.901	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 103.800%	
43) trans-1,3-Dichloroprop...	8.945	79	41834	57.029	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 114.060%	
47) 2-Hexanone-d5	9.378	63	82895	109.003	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 109.000%	
56) 1,1,2,2-Tetrachloroeth...	11.189	84	114102	55.294	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 110.580%	
66) 1,2-Dichlorobenzene-d4	12.317	152	87794	58.555	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 117.120%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030824\
Data File : VX040614.D
Acq On : 07 Mar 2024 09:48
Operator : JC/MD
Sample : VX0308MBL01
Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK662

Quant Time: Mar 08 01:28:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML030724WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 08 01:27:40 2024
Response via : Initial Calibration

