

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030920\
 Data File : VX015137.D
 Acq On : 09 Mar 2020 13:05
 Operator : JC/SP
 Sample : L1842-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE126D3-20200303

Quant Time: Mar 09 14:52:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	336578	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	547788	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	527903	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	263780	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	203279	50.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.82%	
35) Dibromofluoromethane	5.49	113	171372	49.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.18%	
50) Toluene-d8	8.71	98	665406	50.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.24%	
62) 4-Bromofluorobenzene	11.13	95	245650	52.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.78%	
Target Compounds						
44) Trichloroethene	7.21	130	10329	2.469	ug/l	91
64) Tetrachloroethene	9.33	164	6901	1.581	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030920\
Data File : VX015137.D
Acq On : 09 Mar 2020 13:05
Operator : JC/SP
Sample : L1842-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE126D3-20200303

Quant Time: Mar 09 14:52:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

