

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030920\
 Data File : VX015139.D
 Acq On : 09 Mar 2020 13:51
 Operator : JC/SP
 Sample : L1842-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE117D2-20200303

Quant Time: Mar 09 15:24:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	340886	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	542420	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	521249	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	264889	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	206633	51.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.18%	
35) Dibromofluoromethane	5.49	113	171476	50.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.22%	
50) Toluene-d8	8.71	98	657869	50.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.08%	
62) 4-Bromofluorobenzene	11.13	95	238165	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030920\
Data File : VX015139.D
Acq On : 09 Mar 2020 13:51
Operator : JC/SP
Sample : L1842-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE117D2-20200303

Quant Time: Mar 09 15:24:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

