

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030920\
 Data File : VX015142.D
 Acq On : 09 Mar 2020 14:58
 Operator : JC/SP
 Sample : L1842-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE126D2-20200303

Quant Time: Mar 09 15:38:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	340915	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	551294	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	530718	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	275952	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	208545	51.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.12%	
35) Dibromofluoromethane	5.49	113	172440	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
50) Toluene-d8	8.71	98	681934	51.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.10%	
62) 4-Bromofluorobenzene	11.13	95	247522	52.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.90%	
Target Compounds						
12) 1,1-Dichloroethene	2.37	96	3502	1.057	ug/l	96
16) Acetone	2.44	43	6572	4.012	ug/l	100
27) cis-1,2-Dichloroethene	4.59	96	5283	1.331	ug/l	89
44) Trichloroethene	7.21	130	1912154	454.087	ug/l	98
64) Tetrachloroethene	9.33	164	3618	0.825	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030920\
Data File : VX015142.D
Acq On : 09 Mar 2020 14:58
Operator : JC/SP
Sample : L1842-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE126D2-20200303

Quant Time: Mar 09 15:38:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

