

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030920\
 Data File : VX015150.D
 Acq On : 09 Mar 2020 17:59
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 10 11:43:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	323206	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	525166	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	514887	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	262021	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	200465	52.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.56%	
35) Dibromofluoromethane	5.49	113	166092	50.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.26%	
50) Toluene-d8	8.71	98	645044	51.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.36%	
62) 4-Bromofluorobenzene	11.14	95	235115	52.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.60%	
Target Compounds						
44) Trichloroethene	7.21	130	35667	8.891	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030920\
Data File : VX015150.D
Acq On : 09 Mar 2020 17:59
Operator : JC/SP
Sample : IBLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 10 11:43:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

