

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX030921\  
 Data File : VX021109.D  
 Acq On : 09 Mar 2021 15:29  
 Operator : JC/MD  
 Sample : M1596-18 10X  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 INTERCEPTOR-20210305

Quant Time: Mar 10 00:31:39 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X022421W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Feb 24 13:50:55 2021  
 Response via : Initial Calibration

| Compound                     | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|------------------------------|--------|-------|----------|----------|-------|----------|
| -----                        |        |       |          |          |       |          |
| Internal Standards           |        |       |          |          |       |          |
| 1) Pentafluorobenzene        | 5.635  | 168   | 202058   | 50.00    | ug/l  | 0.01     |
| 34) 1,4-Difluorobenzene      | 6.835  | 114   | 376252   | 50.00    | ug/l  | 0.00     |
| 63) Chlorobenzene-d5         | 10.094 | 117   | 380707   | 50.00    | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.059 | 152   | 175952   | 50.00    | ug/l  | 0.00     |
|                              |        |       |          |          |       |          |
| System Monitoring Compounds  |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4    | 6.035  | 65    | 138013   | 50.57    | ug/l  | 0.00     |
| Spiked Amount                | 50.000 | Range | 61 - 141 | Recovery | =     | 101.14%  |
| 35) Dibromofluoromethane     | 5.470  | 113   | 118956   | 52.04    | ug/l  | 0.00     |
| Spiked Amount                | 50.000 | Range | 69 - 133 | Recovery | =     | 104.08%  |
| 50) Toluene-d8               | 8.694  | 98    | 477569   | 53.58    | ug/l  | 0.00     |
| Spiked Amount                | 50.000 | Range | 65 - 126 | Recovery | =     | 107.16%  |
| 62) 4-Bromofluorobenzene     | 11.118 | 95    | 184656   | 56.03    | ug/l  | 0.00     |
| Spiked Amount                | 50.000 | Range | 58 - 135 | Recovery | =     | 112.06%  |
|                              |        |       |          |          |       |          |
| Target Compounds             |        |       |          |          |       | Qvalue   |
| 2) Dichlorodifluoromethane   | 1.187  | 85    | 1827     | 0.84     | ug/l  | 99       |
| 4) Vinyl Chloride            | 1.393  | 62    | 1791     | 0.64     | ug/l  | 97       |
| 9) 1,1,2-Trichlorotrifluo... | 2.370  | 101   | 2016     | 0.93     | ug/l  | 96       |
| 12) 1,1-Dichloroethene       | 2.358  | 96    | 5000     | 2.21     | ug/l  | 91       |
| 21) trans-1,2-Dichloroethene | 3.146  | 96    | 5628     | 2.22     | ug/l  | 98       |
| 27) cis-1,2-Dichloroethene   | 4.570  | 96    | 209679   | 71.64    | ug/l  | 89       |
| 44) Trichloroethene          | 7.188  | 130   | 1080281  | 383.30   | ug/l  | 98       |
| -----                        |        |       |          |          |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX030921\  
Data File : VX021109.D  
Acq On : 09 Mar 2021 15:29  
Operator : JC/MD  
Sample : M1596-18 10X  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
INTERCEPTOR-20210305

Quant Time: Mar 10 00:31:39 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X022421W.M  
Quant Title : SW846 8260  
QLast Update : Wed Feb 24 13:50:55 2021  
Response via : Initial Calibration

